



EXHIBIT B

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 1 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

Revision History:


Revision	Comments
1.0	Initial release for routing

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 2 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

1	INTRODUCTION	3
1.1	Purpose.....	3
1.2	Scope	3
1.3	Document Overview	3
2	REFERENCES AND DEFINITIONS	4
2.1	Internal Medtronic References.....	4
	Note: Document revisions referenced in DVT Plan.	4
2.2	External References	4
2.3	Definitions, Acronyms, and Abbreviations	4
3	TEST RESULTS SUMMARY	5
3.1	Test Paths	6
4	ELECTRICAL TESTS	7
4.1	Power Source Tests	7
4.1.1	Current Drain Test	7
4.1.2	Supply Voltage Range Test.....	9
4.2	Input/Output Connections Tests	11
4.2.1	Keypad Interface Test.....	11
4.2.2	Display Interface Test	12
4.2.3	External Antenna Interface Test.....	14
4.2.4	Infrared Port Interface Test	16
4.2.5	Audio Transducer Test.....	17
4.2.6	Manufacturing/Test Interface Test.....	18
4.3	Internal Resources Tests	19
4.3.1	Memory Test	19
4.3.2	Real-Time Clock Backup Test.....	20
4.3.3	Real-Time Clock Accuracy Test.....	21
4.3.4	A/D Measurements Test	22
4.3.5	D/A Control Voltages Test.....	24
4.4	Transmit Telemetry (Downlink) Tests	25
4.4.1	Magnetic Field Intensity Test.....	25
4.4.2	Burst Characteristics Test.....	27
4.5	Receive Telemetry (Uplink) Tests	30
4.5.1	Detection Threshold Test.....	30
4.5.2	Detection Margin Test.....	32
4.5.3	Noise Immunity Test	34
4.5.4	Signal Distortion Test.....	37
4.5.5	Turnaround Time Test	40
4.5.6	Hold Drift Test.....	41
4.5.7	New-Battery FET Test	43
4.6	Telemetry Performance Tests.....	44
4.6.1	44
5	COMPLETION.....	49

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 3 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

1 INTRODUCTION

This document is the electrical Design Verification Test (DVT) Report for the 37741 Patient Programmer Platform.

1.1 Purpose

The purpose of this report is to document the results of test plan .

1.2 Scope


This report applies only to design verification testing of the 37741 Patient Programmer Platform.

1.3 Document Overview

This document is organized as follows:

- Section 2 contains references and definitions.
- Section 3 contains a table with the list of tests, software revisions, sample sizes, and test results.
- Section 4 contains the results of the electrical design verification tests.

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 4 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

2 REFERENCES AND DEFINITIONS

This section identifies internal and external reference documents that augment the information provided in this document. It also defines terms, acronyms, and abbreviations used within the document.

2.1 Internal Medtronic References

Number	Name
120275	
215387	
288117-70040	
288117-70044	
288117-70029	
503011001	
288117-70200	

Note: Document revisions referenced in DVT Plan.

2.2 External References

Reference the PEM Electrical Specification for external specification standards.

2.3 Definitions, Acronyms, and Abbreviations

ARB: Arbitrary Waveform Generator

ARB equipment: One or more arbitrary waveform generators, used alone or in conjunction to generate sophisticated waveforms.

DUT: Device Under Test

DVT: Design Verification Test

DVT Console: The test console needed to perform the tests specified herein.

ES: Electrical Specification #120275


GPB: General Purpose Interface Bus

PEM: Patient Electronic Module

PP: Patient Programmer

POR: Power On Reset

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 5 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

3 Test Results Summary


Table 1 summarizes the results of all electrical design verification testing. Section 4 details each test setup, criteria, and results.

- Test data is stored as 288117-70200.
- Table 1 indicates test name, sample size, DUT software revision, Test Script Software revision, test path, and results.
- Test paths are shown in section 3.1.

Table 1

Test Name	Sample Size	DUT Software Revision	Script Software Test Revision	Test Path	Results
	22				PASS
	22				PASS
	22				PASS
	22				PASS
	22				PASS
	22				PASS
	22				PASS
	22				PASS
	22				PASS
	22				PASS
	22				PASS
	22				PASS
	22				PASS
	22				PASS
	22				PASS
	22				PASS
	22				PASS
	22				PASS
	22				PASS
	22				PASS
	22				PASS
	22				PASS
	22				PASS
	1				PASS

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 6 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

3.1 Test Paths





EXHIBIT B (cont.)

	Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 7 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report					

4 ELECTRICAL TESTS

This section specifies electrical tests performed on the 37741 Patient Programmer Platform.

4.1 Power Source Tests

4.1.1 Current Drain Test

4.1.1.1 Objective

To verify the current drain meets the requirements specified in the *Power Source* section of the PEM Electrical Specification.

4.1.1.2 Method and Equipment

4.1.1.3 Test Cases

There are _ test cases for transmit using all combinations of test values below:

Parameter	Test Values	Units

The

There are test cases using all combinations of test values below:

Parameter	Test Values	Units


There are test cases using two combinations of test values below:

Parameter	Test Values	Units

There are total test cases.

4.1.1.4 Acceptance Criteria

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 8 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

Operating Condition (Ref.)	Antenna	Duty Cycle (%)	Current Drain (mA) MAX		
			V	V	V
Row A	INT				
Row B	INT				
Row C	INT				
Row D	INT				
Row E	INT				
Row F	INT				
Row G	INT				
Row H	EXT				
Row I	INT				
Row J	INT				

Note 1:

4.1.1.5 Test Setup

- 1.
- 2.
- 3.
- 4.


4.1.1.6 Test Procedure

- 1.
- 2.

3.

4.

EXHIBIT B (cont.)

	Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 9 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report					

4.1.1.7 RESULTS **PASS**

All devices met the acceptance criteria.

Operating Condition	Current Drain (mA) MAX																	
Row	Spec	Min	Max	Mean	Std Dev		Spec	Min	Max	Mean	Std Dev		Spec	Min	Max	Mean	Std Dev	
A																		
B																		
C																		
D																		
E																		
F																		
G																		
H																		
I																		
J																		

4.1.2 **Supply Voltage Range Test**


4.1.2.1 Objective

To verify the supply voltage range meets the requirements specified in the *Power Source* section of the PEM Electrical Specification.

4.1.2.2 Method and Equipment

4.1.2.3 Test Cases

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 10 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

Parameter	Test Values	Units

The

There is test case without transmit:

Parameter	Test Values	Units

4.1.2.4 Acceptance Criteria

Operating Condition	Antenna	H-Bridge Drive Duty Cycle (%)	Min operating voltage (V)


4.1.2.5 Test Setup

- 1.
- 2.
- 3.
- 4.

4.1.2.6 Test Procedure

- 1.
- 2.
- 3.

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 11 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

4.1.2.7 RESULTS **PASS**

All devices met the acceptance criteria.

Operating Condition	Antenna	Supply Voltage Range (Volts)				
		Min	Max	Avg	Std Dev	
—	—	—	—	—	—	—
—	—	—	—	—	—	—

4.2 Input/Output Connections Tests

4.2.1 Keypad Interface Test

4.2.1.1 Objective

To verify the keypad interface meets the requirements specified in the *Input/Output Connections* section of the PEM Electrical Specification.

4.2.1.2 Method and Equipment

4.2.1.3 Test Cases

Parameter	Test Values	Units
—	—	—


4.2.1.4 Acceptance Criteria

4.2.1.5 Test Setup

- 1.
- 2.
- 3.

4.2.1.6 Test Procedure

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 12 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

3.

4.2.1.7 RESULTS **PASS**

All devices met the acceptance criteria.

Tests	Keypad Interface (pass/fail)		
	Pass	Pass	Pass
	Pass	Pass	Pass

4.2.2 Display Interface Test

4.2.2.1 Objective

To verify the display interface meets the requirements specified in the *Input/Output Connections* section of the PEM Electrical Specification.


4.2.2.2 Method and Equipment

4.2.2.3 Test Cases

There are test cases using combinations of the test values below:

Parameter	Test Values	Units

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 13 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

4.2.2.4 Acceptance Criteria

Test Parameters				Requirements			

4.2.2.5 Test Setup

- 1.
- 2.
- 3.

4.2.2.6 Test Procedure


- 1.
- 2.
- 3.
- 4.

4.2.2.7 RESULTS **PASS**

All devices met the acceptance criteria.

Test	Display Interface (pass/fail)		
	Pass	Pass	Pass
	Pass	Pass	Pass
	Pass	Pass	Pass

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 14 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

4.2.3 External Antenna Interface Test**4.2.3.1 Objective**

To verify the external antenna interface meets the requirements specified in the *Input/Output Connections* section of the PEM Electrical Specification.

4.2.3.2 Method and Equipment**4.2.3.3 Test Cases**

There are test cases using all combinations of test values below:

Parameter	Test Values	Units

4.2.3.4 Acceptance Criteria


- When an external antenna is connected, there should be no downlink from the internal antenna.
- When an external antenna is connected, the uP should detect that the antenna is connected.

External Antenna					Yes/No
	Min	Max	Min	Max	

4.2.3.5 Test Setup

- 1.
- 2.
- 3.
- 4.
- 5.

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 15 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

4.2.3.6 Test Procedure

1.

2.


3.

4.

4.2.3.7 RESULTS **PASS**

All devices met the acceptance criteria.

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 16 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

Test Configuration	Test
	A
	B

External Antenna Interface (A/m)														
Test														
	Min	Max	Mean	Std dev		Min	Max	Mean	Std dev		Min	Max	Mean	Std dev
A														
B														

4.2.4 Infrared Port Interface Test

4.2.4.1 Objective

To verify the infrared port interface meets the requirements specified in the *Input/Output Connections* section of the PEM Electrical Specification. [PTPROG_PEMT-0006:1]

4.2.4.2 Method and Equipment

4.2.4.3 Test Cases

There are test cases using all combinations of test values below:

Parameter	Test Values	Units


4.2.4.4 Acceptance Criteria

All	All	None

4.2.4.5 Test Setup

- 1.
- 2.

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 17 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

3.

4.2.4.6 Test Procedure

1.

2.

3.

4.

4.2.4.7 RESULTS **PASS**

All devices met the acceptance criteria.

Voltage (V)	Infrared (pass/fail)								
	Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass
	Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass
	Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass

4.2.5 **Audio Transducer Test**

4.2.5.1 Objective


To verify the audio transducer meets the requirements specified in the *Input/Output Connections* section of the PEM Electrical Specification.

4.2.5.2 Method and Equipment

4.2.5.3 Test Cases

There are test cases using all combinations of test values below:

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 18 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

Parameter	Test Values	Units

4.2.5.4 Acceptance Criteria

4.2.5.5 Test Setup

- 1.
- 2.
- 3.
- 4.
- 5.

4.2.5.6 Test Procedure

- 1.
- 2.
- 3.
- 4.

4.2.5.7 RESULTS **PASS**


All devices met the acceptance criteria.

Audio Transducer (dB SPL)														
Min	Max	Mean	Std dev		Min	Max	Mean	Std dev		Min	Max	Mean	Std dev	

4.2.6 **Manufacturing/Test Interface Test**

Manufacturing requirements defined in Test Specification, Patient Programmer, 215387.

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 19 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

4.3 Internal Resources Tests**4.3.1 Memory Test****4.3.1.1 Objective**

To verify the internal memory resources meet the requirements specified in the *Internal Resources* section of the PEM Electrical Specification.

4.3.1.2 Method and Equipment**4.3.1.3 Test Cases**

There are test cases using all combinations of test values below:

Parameter	Test Values	Units

4.3.1.4 Acceptance Criteria

All	Pass

4.3.1.5 Test Setup

1.

2.

3.

4.3.1.6 Test Procedure

1.

2.

3.


4.

4.3.1.7 RESULTS PASS

All devices met the acceptance criteria.

Test	Memory (pass/fail)		
	Pass	Pass	Pass
	Pass	Pass	Pass
	Pass	Pass	Pass

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 20 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

4.3.2 Real-Time Clock Backup Test

4.3.2.1 Objective

To verify the real-time clock backup meets the requirements specified in the *Internal Resources* section of the PEM Electrical Specification.

4.3.2.2 Method and Equipment

4.3.2.3 Test Cases

There is one test case below:

Parameter	Test Value	Units

4.3.2.4 Acceptance Criteria

Test Case	Min Time w/o power (min)

4.3.2.5 Test Setup

- 1.
- 2.
- 3.


4.3.2.6 Test Procedure

- 1.
- 2.
- 3.
- 4.
- 5.

4.3.2.7 RESULTS **PASS**

All devices met the acceptance criteria.

EXHIBIT B (cont.)

	Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 21 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report					

Test	Real-Time Backup (pass/fail)		
	Pass	Pass	Pass

4.3.3 Real-Time Clock Accuracy Test

4.3.3.1 Objective

To verify the real-time clock accuracy meets the requirements specified in the *Internal Resources* section of the PEM Electrical Specification.


4.3.3.2 Method and Equipment

4.3.3.3 Test Cases

There are test cases (actually measurement points) using all combinations of test values below:

Parameter	Test Value	Units

4.3.3.4 Acceptance Criteria



4.3.3.5 Test Setup

- 1.
- 2.


4.3.3.6 Test Procedure

- 1.
- 2.

4.3.3.7 RESULTS PASS

All devices met the acceptance criteria.

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 22 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

	Real Time Clock Accuracy (seconds)				

4.3.4 A/D Measurements Test

4.3.4.1 Objective

To verify the A/D measurement accuracy meets the requirements specified in the *Internal Resources* section of the PEM Electrical Specification.

4.3.4.2 Method and Equipment

4.3.4.3 Test Cases

There are test cases using the test values below:

Parameter	Test Values	Units


4.3.4.4 Acceptance Criteria

A/D Voltage	Test Value	Max Error (%)

4.3.4.5 Test Setup

- 1.
- 2.
- 3.
- 4.

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 23 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

5.

4.3.4.6 Test Procedure

1.

2.


3.

4.

4.3.4.7 RESULTS **PASS**

All devices met the acceptance criteria.

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 24 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

[illegible]

4.3.5 D/A Control Voltages Test

4.3.5.1 Objective


To verify the D/A accuracy meets the requirements specified in the *Internal Resources* section of the PEM Electrical Specification.

4.3.5.2 Method and Equipment

4.3.5.3 Test Cases

There are test cases using all combinations of test values below:

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 25 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

Parameter	Test Value	Units
-----------	------------	-------

4.3.5.4 Acceptance Criteria

D/A Voltage	Measurement point	Max % Error

4.3.5.5 Test Setup

- 1.
- 2.
- 3.
- 4.

4.3.5.6 Test Procedure

- 1.
- 2.
- 3.

4.3.5.7 RESULTS PASS

All devices met the acceptance criteria.

[illegible]


4.4 Transmit Telemetry (Downlink) Tests

4.4.1 Magnetic Field Intensity Test

4.4.1.1 Objective

To verify downlink magnetic field intensity meets the requirements specified in the *Transmit Telemetry (Downlink)* section of the PEM Electrical Specification.


EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 26 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				


4.4.1.2 Method and Equipment

4.4.1.3 Test Cases

There are test cases at kHz using all combinations of test values below:



4.4.1.4 Acceptance Criteria




4.4.1.5 Test Setup

- 1.
- 2.
- 3.
- 4.
- 5.

4.4.1.6 Test Procedure

- 1.
- 2.
- 3.
- 4.

EXHIBIT B (cont.)

	Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 27 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report					

5.

4.4.1.7 RESULTS PASS

All devices met the acceptance criteria.

	Magnetic Field Intensity (A/m)

4.4.2 Burst Characteristics Test**4.4.2.1 Objective**


To verify downlink burst characteristics of width, rise time, fall time, frequency, and overshoot meet the requirements specified in the *Transmit Telemetry (Downlink)* section of the PEM Electrical Specification.

4.4.2.2 Method and Equipment**4.4.2.3 Test Cases**

There are test cases using all combinations of test values below:

Parameter	Test Values	Units

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 28 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

4.4.2.4 Acceptance Criteria

--

4.4.2.5 Test Setup

- 1.
- 2.
- 3.
- 4.
- 5.

4.4.2.6 Test Procedure

- 1.
- 2.
- 3.
- 4.
- 5.


4.4.2.7 RESULTS **PASS**

All devices met the acceptance criteria.

Sht
29 of 49

[illegible]

EXHIBIT B (cont.)

 Medtronic	Neurological	Document Number 288117-70205	Rev/Version 1.0	Sht 30 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

4.5 Receive Telemetry (Uplink) Tests


4.5.1 Detection Threshold Test

4.5.1.1 Objective

To verify uplink detection threshold (i.e. receiver sensitivity) meets the requirements specified in the *Receive Telemetry (Uplink)* section of the PEM Electrical Specification.

4.5.1.2 Method and Equipment

Figure 1: Example Uplink Detection Threshold Test Waveforms




4.5.1.3 Test Cases

There are test cases using all combinations of test values below:

Parameter	Test Values	Units

The supply voltage is 2.5 V.

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 31 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

4.5.1.4 Acceptance Criteria

Antenna	Telemetry Type	Detection Onset (Uplink dB)	Detection Threshold (Uplink dB)	Maximum Input Level (Uplink dB)
		Max	Max	Max

4.5.1.5 Test Setup

- 1.
- 2.
- 3.
- 4.
- 5.

4.5.1.6 Test Procedure


- 1.
- 2.
- 3.
- 4.

4.5.1.7 RESULTS **PASS**

All devices met the acceptance criteria.

Antenna	Telemetry	Detection Threshold (dB)														
		Min	Max	Mean	Std dev		Min	Max	Mean	Std dev		Min	Max	Mean	Std dev	

EXHIBIT B (cont.)

	Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 32 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report					

Antenna	Telemetry	Maximum Input Level (pass/fail)		
		Pass	Pass	Pass
		Pass	Pass	Pass
		Pass	Pass	Pass
		Pass	Pass	Pass
		Pass	Pass	Pass
		Pass	Pass	Pass


4.5.2 Detection Margin Test

4.5.2.1 Objective

To verify uplink detection margin meets the requirements specified in the *Receive Telemetry (Uplink)* section of the PEM Electrical Specification.

4.5.2.2 Method and Equipment


Figure 2: Example Uplink Detection Margin Test Waveforms



4.5.2.3 Test Cases

There are test cases using all combinations of test values below:

EXHIBIT B (cont.)

	Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 33 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report					

Parameter	Test Values	Units

4.5.2.4 Acceptance Criteria

Telemetry Type	Data Bursts	Amplitude A1	Antenna	Detection Margin (Uplink dB)	
				Min	Max

4.5.2.5 Test Setup

- 1.
- 2.
- 3.
- 4.
- 5.

4.5.2.6 Test Procedure

- 1.
- 2.
- 3.
- 4.

4.5.2.7 RESULTS **PASS**

All devices met the acceptance criteria.

4.5.3.2 Method and Equipment

EXHIBIT B (cont.)



	Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 35 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report					

Figure 3: Example Uplink Noise Immunity Test Waveforms




4.5.3.3 Test Cases

There are test cases using all combinations of test values below:

Parameter	Test Values	Units

EXHIBIT B (cont.)

	Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 36 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report					

4.5.3.4 Acceptance Criteria

Telemetry Type	Uplink Level A1 (dB)	Antenna	Min E-Noise Immunity (dB)	Min H-Noise Immunity (dB)

4.5.3.5 Test Setup

- 1.
- 2.
- 3.
- 4.
- 5.
- 6.


4.5.3.6 Test Procedure

- 1.
- 2.
- 3.
- 4.
- 5.

4.5.3.7 RESULTS **PASS**

All devices met the acceptance criteria.

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 37 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

[illegible]


4.5.4 Signal Distortion Test

4.5.4.1 Objective

To verify uplink signal distortion meets the requirements specified in the *Receive Telemetry (Uplink)* section of the PEM Electrical Specification.

4.5.4.2 Method and Equipment

EXHIBIT B (cont.)

	Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 38 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report					

4.5.4.3 Test Cases

Parameter	Test Values	Units

There are test cases for Tel A, and test cases for Tel N.

4.5.4.4 Acceptance Criteria

Telemetry Type	Uplink Level A1 (dB)	Antenna	Interval Distortion (µS)	Active/Idle Distortion (µS)

4.5.4.5 Test Setup

- 1.
- 2.
- 3.
- 4.

5.

4.5.4.6 Test Procedure


- 1.
- 2.
- 3.

4.

4.5.4.7 RESULTS **PASS**


All devices met the acceptance criteria.

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 39 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

[illegible][illegible]

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 40 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

[illegible]

4.5.5 Turnaround Time Test

4.5.5.1 Objective

To verify uplink turnaround time meets the requirements specified in the *Receive Telemetry (Uplink)* section of the PEM Electrical Specification.


4.5.5.2 Method and Equipment

4.5.5.3 Test Cases

There are $2^4 = 16$ test cases using all combinations of test values below:

Parameter	Test Values	Units

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 41 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

4.5.5.4 Acceptance Criteria

Supply Voltage	H-Bridge Drive Duty Cycle	Turnaround Time (mS)

4.5.5.5 Test Setup

- 1.
- 2.
- 3.

4.5.5.6 Test Procedure

- 1.
- 2.
- 3.
- 4.

4.5.5.7 RESULTS **PASS**

All devices met the acceptance criteria.

	Turnaround Time (pass/fail)
Test	


4.5.6 **Hold Drift Test**

4.5.6.1 Objective

To verify the hold drift meets the requirements specified in the *Receive Telemetry (Uplink)* section of the PEM Electrical Specification.

4.5.6.2 Method and Equipment

EXHIBIT B (cont.)

	Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 42 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report					

4.5.6.3 Test Cases

There is test case:

Parameter	Uplink Level	Units

4.5.6.4 Acceptance Criteria

Time after hold circuit enabled	Max Hold Drift


4.5.6.5 Test Setup

- 1.
- 2.
- 3.

4.5.6.6 Test Procedure

- 1.
- 2.
- 3.
- 4.
- 5.
- 6.
- 7.
- 8.
- 9.
- 10

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 43 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

4.5.6.7 RESULTS **PASS**

All devices met the acceptance criteria.

Hold Drift (mV)
<div></div>

4.5.7 **New-Battery FET Test**

4.5.7.1 Objective

To verify that enabling the new-battery FET circuit reduces the receiver noise floor (ambient RF energy detected by the receiver circuit) when new batteries are used.

4.5.7.2 Method and Equipment

4.5.7.3 Test Cases

There is test case:

Parameter	Uplink Level	Units


4.5.7.4 Acceptance Criteria

Supply Voltage	New-Battery FET	RSSI Peak

4.5.7.5 Test Setup

- 1.
- 2.
- 3.

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 44 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

4.5.7.6 Test Procedure

- 1.
- 2.
- 3.
- 4.
- 5.
- 6.
- 7.

4.5.7.7 RESULTS **PASS**

All devices met the acceptance criteria.

New-Battery FET (mV)													
Min	Max	Mean	Std dev		Min	Max	Mean	Std dev		Min	Max	Mean	Std dev

4.6 Telemetry Performance Tests


4.6.1 Telemetry Map Area at a Fixed Distance Test

4.6.1.1 Objective

To verify telemetry performance in terms of map area at a fixed distance meets the requirements specified in the *Telemetry Performance* section of the PEM Electrical Specification.

4.6.1.2 Method and Equipment

EXHIBIT B (cont.)

	Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 45 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report					

4.6.1.3 Test Cases

Parameter	Test Values	Units

There are test cases.

4.6.1.4 Acceptance Criteria

IPG	Antenna	Map Area @ 5cm

4.6.1.5 Test Setup


- 1.
- 2.

4.6.1.6 Test Procedure


- 1.
- 2.
- 3.
- 4.
- 5.
- 6.
- 7.
- 8.

4.6.1.7 RESULTS **PASS**

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 46 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

4.6.1.7.1 Internal Antenna Map @



4.6.1.7.2 Internal Antenna @







EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 47 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

4.6.1.7.3 External Antenna Map @




4.6.1.7.4 External Antenna @




4.6.1.7.5 Photo of test fixture showing

EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 48 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				


in this photo.



4.6.1.7.6 Photo of



EXHIBIT B (cont.)

 Medtronic	<i>Neurological</i>	Document Number 288117-70205	Rev/Version 1.0	Sht 49 of 49
Title: Neuro Patient Programmer Platform Electrical DVT Report				

5 COMPLETION

This paragraph concludes this test specification.